

CERTIFICATE OF ANALYSIS

Product : **ATRAZINE**
Product number : **A0156**
Batch number : **010948**
Date of production : **07-06-2013**
Date of analysis : **24-06-2013**
Expiry date : **30-06-2018**

Packaging and storage : store in a well-closed container at room temperature

	SPECIFICATION	RESULT	UNITS	REMARKS
Identity	conform	complies		NMR
Melting range		175.3 – 177.6	°C	
Water		0.19	%	Karl Fischer
Assay		99.1	%	HPLC


This certificate of analysis is conform the results of the manufacturer.

Conclusion: approved

Reviewed by:

drs. M. van Wissen, pharmacist
Department of Quality Control Duchefa B.V.

Date: 07-01-2015



The buyer is responsible to test each batch to ensure the product is suitable for its specific purpose.