

CERTIFICATE OF ANALYSIS

Product : **AMMONIUM CHLORIDE**
 Product no. : A0528
 Batchnumber : 012414
 Date of production : xx-06-2016
 Date of expiry : 30-06-2021
 :
 Packaging and storage : in a well-closed container

	SPECIFICATION	RESULT	UNITS	REMARKS
Identity:				
- A (chloride)	conform	complies		
- B (ammonium)	conform	complies		
Appearance of solution	clear and colourless	complies		10% in H ₂ O
Acidity or alkalinity	conform	complies		
Bromides and iodides	conform	complies		
Sulphates	≤ 150	< 100	ppm	
Calcium	≤ 200	< 10	ppm	
Iron	≤ 20	< 1.0	ppm	
Heavy metals	≤ 10	< 10	ppm	
Loss on drying	≤ 1.0	0.14	%	105 °C
Sulphated ash	≤ 0.1	< 0.10	%	
Assay	99.0 – 100.5	99.2	%	


This certificate of analysis is conform the results of the manufacturer.

Conclusion: approved

Released by:

drs. M. van Wissen, pharmacist
 Department of Quality Control Duchefa B.V.

Date: 28-09-2016



The buyer is responsible to test each batch to ensure the product is suitable for its specific purpose.