

CERTIFICATE OF ANALYSIS

Product : **p-AMINO BENZOIC ACID**
Product number : A0601
Batch number : 018192
Date of analysis : 20-09-2022
Expiry date : 30-09-2025

Packaging and storage : store in a well-closed container at room temperature


	SPECIFICATION	RESULT	UNITS	REMARKS
Appearance	(almost) white, crystalline powder or crystals	complies		
Identity	conform	complies		IR
Melting range (upper value)	186	186	°C	
Assay	98.5 – 101.5	99.6	%	acidimetric

This certificate of analysis is conform the results of the manufacturer.

Conclusion: approved
Released by:

drs. M. van Wissen, pharmacist
Department of Quality Control Duchefa B.V.

Date: 18-12-2023



The buyer is responsible to test each batch to ensure the product is suitable for their specific purpose.

Duchefa Biochemie B.V.